

Certificate of Analysis

100 bp DNA Ladder RTU

Part No./Catalog No. DM001-R500

Lot Number: DM17660052

Size: 50µg/500µl

5 ul per well for 100 gel applications

Product Qualification

Specifications: The product must show 11 dsDNA bands when separated by electrophoresis on a 1.7% agarose gel with 1x TBE buffer.

Results: Meet specifications.

Specification: Any contaminant band must be less than 5 % in intensity to the nearest major band shown by Novel Juice staining.

Result: Meets specification.

All products are for research use only.

Caution: Not intended for human or animal diagnostic or therapeutic uses.



5ul DM001-R500, with 1ul Novel Juice (LD001-1000)

www.genedirex.com
info@genedirex.com

GeneDireX[®]