

Certificate of Analysis



PiNK Plus Prestained Ladder

Part No./Catalog No. PM005-0500

Lot Number: 70104257

Size: 500 ul

Product Qualification

Specifications: The product must show 11 protein bands when separated by SDS PAGE electrophoresis on an OnePAGE Sonic Gel with Tris-Glycine buffer, 12% Bis-Tris Gel with MOPS buffer, or 4-12% Bis-Tris Gel with MES buffer.

Results: Meet specifications.

Specification: Any contaminant band must be less than 10% in intensity to the nearest major band.

Result: Meets specification.

Specification: Pre-stained protein band from lot being tested must migrate within 1/2 bandwidth from prior lot.

Result: Meets specification.

All products are for research use only.

Caution: Not intended for human or animal diagnostic or therapeutic uses.



5ul, OnePAGE Sonic Gel with Tris/ Glycine/ SDS buffer